

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: WEST04

Lab Code: CHEM Case No.: P5117 SAS No.: P5117 SDG No.: P5117

Instrument ID: MSVOA_N Calibration Date/Time: 12/09/2024 16:14

Lab File ID: VN085153.D Init. Calib. Date(s): 11/22/2024 11/22/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:22 13:45

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.589	0.577		-2.04	50
1,1-Dichloroethene	0.531	0.571		7.53	50
2-Butanone	0.339	0.438		29.2	50
Carbon Tetrachloride	0.527	0.629		19.35	50
Chloroform	1.146	1.269		10.73	50
Benzene	1.445	1.610		11.5	50
1,2-Dichloroethane	0.494	0.589		19.23	50
Trichloroethene	0.336	0.391		16.37	50
Tetrachloroethene	0.337	0.394		16.91	50
Chlorobenzene	1.134	1.321	0.3	16.49	50
1,2-Dichloroethane-d4	0.712	0.726		1.97	50
Dibromofluoromethane	0.337	0.362		7.42	50
Toluene-d8	1.208	1.305		8.03	50
4-Bromofluorobenzene	0.452	0.506		11.95	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.