

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5117

Client: Weston Solutions

Analytical Method: SW8260-Low

Datafile : VN085144.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1209WBS01	Vinyl chloride	20	15.7	ug/L	79			58	137	
	1,1-Dichloroethene	20	16.8	ug/L	84			71	131	
	2-Butanone	100	98.2	ug/L	98			56	143	
	Carbon Tetrachloride	20	19.8	ug/L	99			72	136	
	Chloroform	20	17.6	ug/L	88			79	124	
	Benzene	20	18.6	ug/L	93			79	120	
	1,2-Dichloroethane	20	18.8	ug/L	94			73	128	
	Trichloroethene	20	19.0	ug/L	95			79	123	
	Tetrachloroethene	20	20.8	ug/L	104			74	129	
	Chlorobenzene	20	18.3	ug/L	92			82	118	