

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5117

Client: Weston Solutions

Analytical Method: SW8260-Low

Datafile : VN085145.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1209WBSD01	Vinyl chloride	20	16.7	ug/L	84	6		58	137	20
	1,1-Dichloroethene	20	18.2	ug/L	91	8		71	131	20
	2-Butanone	100	120	ug/L	120	20		56	143	20
	Carbon Tetrachloride	20	21.0	ug/L	105	6		72	136	20
	Chloroform	20	19.5	ug/L	98	11		79	124	20
	Benzene	20	20.0	ug/L	100	7		79	120	20
	1,2-Dichloroethane	20	20.6	ug/L	103	9		73	128	20
	Trichloroethene	20	21.0	ug/L	105	10		79	123	20
	Tetrachloroethene	20	20.7	ug/L	104	0		74	129	20
	Chlorobenzene	20	19.7	ug/L	99	7		82	118	20