

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WEST04
 Lab Code: CHEM Case No.: P5117 SAS No.: P5117 SDG NO.: P5117
 Lab File ID: VN085142.D Date Analyzed: 12/09/2024
 Instrument ID: MSVOA_N Time Analyzed: 10:32
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	120023	13.788				
UPPER LIMIT	240046	14.288				
LOWER LIMIT	60011.5	13.288				
EPA SAMPLE NO.						
TAPIAL2-IDW-SOIL-120424-00-T2	85732	13.79				
VN1209WBL01	78404	13.79				
VN1209WBS01	104600	13.79				
VN1209WBSD01	101401	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.