

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: P5120

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB133773

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.097	0.099	98	85-115	12/05/2024
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	12/05/2024
Sample ID: CCV2 Reactive Cyanide	mg/L	0.23	0.25	92	90-110	12/05/2024
Sample ID: CCV3 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	12/05/2024