

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P5123 SAS No.: P5123 SDG NO.: P5123

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL01 Date Analyzed: 12/06/2024

Lab Sample No.: AR1660CCC500 Data File : PP068866.D Time Analyzed: 21:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.923	5.821	6.021	459.220	500.000	-8.2
Aroclor-1016-2	5.945	5.844	6.044	459.770	500.000	-8.0
Aroclor-1016-3	6.008	5.908	6.108	497.910	500.000	-0.4
Aroclor-1016-4	6.107	6.005	6.205	505.210	500.000	1.0
Aroclor-1016-5	6.401	6.299	6.499	413.540	500.000	-17.3
Aroclor-1260-1	7.524	7.423	7.623	411.730	500.000	-17.7
Aroclor-1260-2	7.778	7.676	7.876	416.730	500.000	-16.7
Aroclor-1260-3	8.140	8.038	8.238	425.060	500.000	-15.0
Aroclor-1260-4	8.378	8.276	8.476	399.870	500.000	-20.0
Aroclor-1260-5	8.715	8.613	8.813	412.920	500.000	-17.4
Decachlorobiphenyl	10.673	10.572	10.772	44.220	50.000	-11.6
Tetrachloro-m-xylene	4.759	4.657	4.857	55.190	50.000	10.4

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P5123 SAS No.: P5123 SDG NO.: P5123

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL01 Date Analyzed: 12/06/2024

Lab Sample No.: AR1660CCC500 Data File : PP068866.D Time Analyzed: 21:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.152	5.058	5.258	462.580	500.000	-7.5
Aroclor-1016-2	5.173	5.079	5.279	497.290	500.000	-0.5
Aroclor-1016-3	5.353	5.259	5.459	481.120	500.000	-3.8
Aroclor-1016-4	5.393	5.298	5.498	496.540	500.000	-0.7
Aroclor-1016-5	5.611	5.517	5.717	476.910	500.000	-4.6
Aroclor-1260-1	6.653	6.560	6.760	472.320	500.000	-5.5
Aroclor-1260-2	6.840	6.747	6.947	466.570	500.000	-6.7
Aroclor-1260-3	6.998	6.905	7.105	470.290	500.000	-5.9
Aroclor-1260-4	7.471	7.378	7.578	489.300	500.000	-2.1
Aroclor-1260-5	7.710	7.617	7.817	475.850	500.000	-4.8
Decachlorobiphenyl	9.200	9.110	9.310	47.100	50.000	-5.8
Tetrachloro-m-xylene	4.048	3.951	4.151	50.570	50.000	1.1