

### CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P5123 SAS No.: P5123 SDG NO.: P5123

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL02 Date Analyzed: 12/07/2024

Lab Sample No.: AR1660CCC500 Data File : PP068875.D Time Analyzed: 03:23

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|--------|-----------|--------|--------------------|-------------------|-------|
|                      |        | FROM      | TO     |                    |                   |       |
| Aroclor-1016-1       | 5.926  | 5.821     | 6.021  | 464.580            | 500.000           | -7.1  |
| Aroclor-1016-2       | 5.948  | 5.844     | 6.044  | 465.370            | 500.000           | -6.9  |
| Aroclor-1016-3       | 6.012  | 5.908     | 6.108  | 492.050            | 500.000           | -1.6  |
| Aroclor-1016-4       | 6.110  | 6.005     | 6.205  | 513.520            | 500.000           | 2.7   |
| Aroclor-1016-5       | 6.404  | 6.299     | 6.499  | 406.820            | 500.000           | -18.6 |
| Aroclor-1260-1       | 7.528  | 7.423     | 7.623  | 406.390            | 500.000           | -18.7 |
| Aroclor-1260-2       | 7.782  | 7.676     | 7.876  | 411.460            | 500.000           | -17.7 |
| Aroclor-1260-3       | 8.143  | 8.038     | 8.238  | 417.190            | 500.000           | -16.6 |
| Aroclor-1260-4       | 8.382  | 8.276     | 8.476  | 395.760            | 500.000           | -20.8 |
| Aroclor-1260-5       | 8.719  | 8.613     | 8.813  | 407.460            | 500.000           | -18.5 |
| Decachlorobiphenyl   | 10.679 | 10.572    | 10.772 | 43.340             | 50.000            | -13.3 |
| Tetrachloro-m-xylene | 4.762  | 4.657     | 4.857  | 57.920             | 50.000            | 15.8  |

### CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: P5123 SAS No.: P5123 SDG NO.: P5123

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL02 Date Analyzed: 12/07/2024

Lab Sample No.: AR1660CCC500 Data File : PP068875.D Time Analyzed: 03:23

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 5.155 | 5.058     | 5.258 | 501.060            | 500.000           | 0.2  |
| Aroclor-1016-2       | 5.174 | 5.079     | 5.279 | 497.030            | 500.000           | -0.6 |
| Aroclor-1016-3       | 5.355 | 5.259     | 5.459 | 493.180            | 500.000           | -1.4 |
| Aroclor-1016-4       | 5.394 | 5.298     | 5.498 | 504.530            | 500.000           | 0.9  |
| Aroclor-1016-5       | 5.612 | 5.517     | 5.717 | 475.770            | 500.000           | -4.8 |
| Aroclor-1260-1       | 6.655 | 6.560     | 6.760 | 467.360            | 500.000           | -6.5 |
| Aroclor-1260-2       | 6.842 | 6.747     | 6.947 | 461.700            | 500.000           | -7.7 |
| Aroclor-1260-3       | 6.999 | 6.905     | 7.105 | 466.300            | 500.000           | -6.7 |
| Aroclor-1260-4       | 7.473 | 7.378     | 7.578 | 483.670            | 500.000           | -3.3 |
| Aroclor-1260-5       | 7.711 | 7.617     | 7.817 | 472.540            | 500.000           | -5.5 |
| Decachlorobiphenyl   | 9.201 | 9.110     | 9.310 | 48.160             | 50.000            | -3.7 |
| Tetrachloro-m-xylene | 4.050 | 3.951     | 4.151 | 51.330             | 50.000            | 2.7  |