

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** P5138
Contract: TULL01 **Lab Code:** CHEM **Case No.:** P5138 **SAS No.:** P5138
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	528	510	104	95 - 105	P	12/06/2024	12:14	LB133790
	Iron	9870	10000	99	95 - 105	P	12/06/2024	12:14	LB133790
	Lead	1000	1000	100	95 - 105	P	12/06/2024	12:14	LB133790

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** P5138
Contract: TULL01 **Lab Code:** CHEM **Case No.:** P5138 **SAS No.:** P5138
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	22.1	20.0	111	80 - 120	P	12/06/2024	12:20	LB133790
	Iron	98.1	100	98	80 - 120	P	12/06/2024	12:20	LB133790
	Lead	10.5	12.0	88	80 - 120	P	12/06/2024	12:20	LB133790

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** P5138
Contract: TULL01 **Lab Code:** CHEM **Case No.:** P5138 **SAS No.:** P5138
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1280	1250	102	90 - 110	P	12/06/2024	13:09	LB133790
	Iron	4940	5000	99	90 - 110	P	12/06/2024	13:09	LB133790
	Lead	5020	5000	100	90 - 110	P	12/06/2024	13:09	LB133790
CCV02	Copper	1250	1250	100	90 - 110	P	12/06/2024	13:38	LB133790
	Iron	4810	5000	96	90 - 110	P	12/06/2024	13:38	LB133790
	Lead	4900	5000	98	90 - 110	P	12/06/2024	13:38	LB133790
CCV03	Copper	1230	1250	99	90 - 110	P	12/06/2024	14:42	LB133790
	Iron	4770	5000	95	90 - 110	P	12/06/2024	14:42	LB133790
	Lead	4880	5000	98	90 - 110	P	12/06/2024	14:42	LB133790
CCV04	Copper	1240	1250	100	90 - 110	P	12/06/2024	15:40	LB133790
	Iron	4880	5000	98	90 - 110	P	12/06/2024	15:40	LB133790
	Lead	4960	5000	99	90 - 110	P	12/06/2024	15:40	LB133790
CCV05	Copper	1230	1250	99	90 - 110	P	12/06/2024	16:35	LB133790
	Iron	4850	5000	97	90 - 110	P	12/06/2024	16:35	LB133790
	Lead	4890	5000	98	90 - 110	P	12/06/2024	16:35	LB133790
CCV06	Copper	1260	1250	101	90 - 110	P	12/06/2024	17:27	LB133790
	Iron	4910	5000	98	90 - 110	P	12/06/2024	17:27	LB133790
	Lead	5020	5000	100	90 - 110	P	12/06/2024	17:27	LB133790
CCV07	Copper	1230	1250	99	90 - 110	P	12/06/2024	18:21	LB133790
	Iron	4760	5000	95	90 - 110	P	12/06/2024	18:21	LB133790
	Lead	4930	5000	99	90 - 110	P	12/06/2024	18:21	LB133790