



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Environmental, Inc

SDG No.: P5138

Contract: TULL01

Lab Code: CHEM

Case No.: P5138

SAS No.: P5138

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Copper	22.0	20.0	110	40 - 160	P	12/06/2024	12:28	LB133790
	Iron	96.6	100	97	40 - 160	P	12/06/2024	12:28	LB133790
	Lead	11.1	12.0	93	40 - 160	P	12/06/2024	12:28	LB133790