



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc **SDG No.:** P5138
Contract: TULL01 **Lab Code:** CHEM **Case No.:** P5138 **SAS No.:** P5138

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Copper	20.0	+/-20.0	U	20.0	P	12/06/2024	12:24	LB133790
	Iron	100	+/-100	U	100	P	12/06/2024	12:24	LB133790
	Lead	12.0	+/-12.0	U	12.0	P	12/06/2024	12:24	LB133790

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc SDG No.: P5138
Contract: TULL01 Lab Code: CHEM Case No.: P5138 SAS No.: P5138

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Copper	20.0	+/-20.0	U	20.0	P	12/06/2024	13:13	LB133790
	Iron	100	+/-100	U	100	P	12/06/2024	13:13	LB133790
	Lead	12.0	+/-12.0	U	12.0	P	12/06/2024	13:13	LB133790
CCB02	Copper	20.0	+/-20.0	U	20.0	P	12/06/2024	13:42	LB133790
	Iron	100	+/-100	U	100	P	12/06/2024	13:42	LB133790
	Lead	12.0	+/-12.0	U	12.0	P	12/06/2024	13:42	LB133790
CCB03	Copper	20.0	+/-20.0	U	20.0	P	12/06/2024	14:46	LB133790
	Iron	100	+/-100	U	100	P	12/06/2024	14:46	LB133790
	Lead	12.0	+/-12.0	U	12.0	P	12/06/2024	14:46	LB133790
CCB04	Copper	20.0	+/-20.0	U	20.0	P	12/06/2024	15:44	LB133790
	Iron	100	+/-100	U	100	P	12/06/2024	15:44	LB133790
	Lead	12.0	+/-12.0	U	12.0	P	12/06/2024	15:44	LB133790
CCB05	Copper	20.0	+/-20.0	U	20.0	P	12/06/2024	16:39	LB133790
	Iron	100	+/-100	U	100	P	12/06/2024	16:39	LB133790
	Lead	12.0	+/-12.0	U	12.0	P	12/06/2024	16:39	LB133790
CCB06	Copper	20.0	+/-20.0	U	20.0	P	12/06/2024	17:31	LB133790
	Iron	100	+/-100	U	100	P	12/06/2024	17:31	LB133790
	Lead	12.0	+/-12.0	U	12.0	P	12/06/2024	17:31	LB133790
CCB07	Copper	20.0	+/-20.0	U	20.0	P	12/06/2024	18:27	LB133790
	Iron	100	+/-100	U	100	P	12/06/2024	18:27	LB133790
	Lead	12.0	+/-12.0	U	12.0	P	12/06/2024	18:27	LB133790