

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>P5138</u>
contract: <u>TULL01</u>	lab code: <u>CHEM</u>	case no.: <u>P5138</u> sas no.: <u>P5138</u>
matrix: <u>Water</u>	sample id: <u>P5138-02</u>	client id: <u>002-35TH-AVE(NOV)MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>P5138-02MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	176		15.5		150	107		P
Iron	ug/L	75 - 125	6940		5210		1500	115		P
Lead	ug/L	75 - 125	515		23.5		500	98		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>P5138</u>
contract: <u>TULL01</u>	lab code: <u>CHEM</u>	case no.: <u>P5138</u> sas no.: <u>P5138</u>
matrix: <u>Water</u>	sample id: <u>P5138-02</u>	client id: <u>002-35TH-AVE(NOV)MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>P5138-02MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	177		15.5		150	108		P
Iron	ug/L	75 - 125	6950		5210		1500	116		P
Lead	ug/L	75 - 125	518		23.5		500	99		P