

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Environmental, Inc **SDG No.:** P5138
Contract: TULL01 **Lab Code:** CHEM **Case No.:** P5138 **SAS No.:** P5138

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165430BS							
Copper	ug/L	150	152		101	85 - 115	P
Iron	ug/L	1500	1420		95	85 - 115	P
Lead	ug/L	500	479		96	85 - 115	P