

Initial and Continuing Calibration Verification

Client: Tully Environmental, Inc

SDG No.: P5139

Project: Transfer Station-SPDES

RunNo.: LB133842

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	12/09/2024
Sample ID: CCV1 Ammonia as N	mg/L	1.1	1	110	90-110	12/09/2024
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	12/09/2024
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	12/09/2024
Sample ID: CCV4 Ammonia as N	mg/L	1	1	100	90-110	12/09/2024

Initial and Continuing Calibration Verification

Client: Tully Environmental, Inc

SDG No.: P5139

Project: Transfer Station-SPDES

RunNo.: LB133842

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------