

Initial and Continuing Calibration Blank Summary

Client: Tully Environmental, Inc

SDG No.: P5139

Project: Transfer Station-SPDES

RunNo.: LB133842

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/09/2024
Sample ID: CCB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/09/2024
Sample ID: CCB2 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/09/2024
Sample ID: CCB3 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/09/2024
Sample ID: CCB4 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/09/2024

Initial and Continuing Calibration Blank Summary

Client: Tully Environmental, Inc

SDG No.: P5139

Project: Transfer Station-SPDES

RunNo.: LB133842

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
---------	-------	--------	----------------------	--------------	-----	-----	------------------