

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc **SDG No.:** P5139
Contract: TULL01 **Lab Code:** CHEM **Case No.:** P5139 **SAS No.:** P5139
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	-0.59	2.0	30	-18	22	12/06/2024	12:50	LB133790
	Iron	101000	101000	100	85600	116500	12/06/2024	12:50	LB133790
	Lead	11.9			-12	12	12/06/2024	12:50	LB133790
ICSAB01	Copper	486	511	95	434	588	12/06/2024	12:55	LB133790
	Iron	99200	99300	100	84400	114500	12/06/2024	12:55	LB133790
	Lead	60.5	49.0	124	37	61	12/06/2024	12:55	LB133790
ICSA	Copper	0.47	2.0	23	-18	22	12/06/2024	13:00	LB133790
	Iron	96500	101000	96	85600	116500	12/06/2024	13:00	LB133790
	Lead	-15.2			-12	12	12/06/2024	13:00	LB133790
ICSAB	Copper	522	511	102	434	588	12/06/2024	13:04	LB133790
	Iron	97400	99300	98	84400	114500	12/06/2024	13:04	LB133790
	Lead	50.0	49.0	102	37	61	12/06/2024	13:04	LB133790