

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TULL01

Lab Code: CHEM Case No.: P5139 SAS No.: P5139 SDG No.: P5139

Instrument ID: MSVOA_N Calibration Date/Time: 12/06/2024 10:26

Lab File ID: VN085118.D Init. Calib. Date(s): 12/06/2024 12/06/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:02 11:37

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Benzene	1.521	1.501		-1.32	
Toluene	1.750	1.713		-2.11	
Ethyl Benzene	1.857	1.739		-6.35	
m/p-Xylenes	0.714	0.713		-0.14	
o-Xylene	0.671	0.648		-3.43	
1,2-Dichloroethane-d4	2.380	2.393		0.55	
Toluene-d8	1.407	1.443		2.56	
4-Bromofluorobenzene	0.496	0.495		-0.2	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.