



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Blank Summary

Client: VERINA CONSULTING GROUP, LLC

SDG No.: P5141

Project: Rotor Clip - PO# 5183.0001

RunNo.: LB133768

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Residual Chlorine	mg/L	< 0.0500	0.0500	U	0.016	0.1	12/05/2024
Sample ID: CCB1 Residual Chlorine	mg/L	< 0.0500	0.0500	U	0.016	0.1	12/05/2024
Sample ID: CCB2 Residual Chlorine	mg/L	< 0.0500	0.0500	U	0.016	0.1	12/05/2024

Initial and Continuing Calibration Blank Summary

Client: VERINA CONSULTING GROUP, LLC

SDG No.: P5141

Project: Rotor Clip - PO# 5183.0001

RunNo.: LB133837

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB COD	mg/L	< 5.0000	5.0000	U	2.35	10	10/14/2024
Sample ID: CCB1 COD	mg/L	< 5.0000	5.0000	U	2.35	10	12/09/2024
Sample ID: CCB2 COD	mg/L	< 5.0000	5.0000	U	2.35	10	12/09/2024
Sample ID: CCB3 COD	mg/L	< 5.0000	5.0000	U	2.35	10	12/09/2024

Initial and Continuing Calibration Blank Summary

Client: VERINA CONSULTING GROUP, LLC

SDG No.: P5141

Project: Rotor Clip - PO# 5183.0001

RunNo.: LB133842

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/09/2024
Sample ID: CCB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/09/2024
Sample ID: CCB2 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/09/2024
Sample ID: CCB3 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/09/2024
Sample ID: CCB4 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/09/2024

Initial and Continuing Calibration Blank Summary

Client: VERINA CONSULTING GROUP, LLC

SDG No.: P5141

Project: Rotor Clip - PO# 5183.0001

RunNo.: LB133847

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	12/09/2024
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	12/09/2024
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	12/09/2024
Sample ID: CCB3 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	12/09/2024