

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	P5141
Project:	Rotor Clip - PO# 5183.0001	Run No.:	LB133768

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133768BS							
Residual Chlorine	mg/L	0.4	0.41		104	1	90-110	12/05/2024

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	P5141
Project:	Rotor Clip - PO# 5183.0001	Run No.:	LB133770

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133770BS							
BOD5	mg/L	198	212		107	1	84.6-115.4	12/05/2024

Laboratory Control Sample Summary

Client: VERINA CONSULTING GROUP, LLC

SDG No.: P5141

Project: Rotor Clip - PO# 5183.0001

Run No.: LB133785

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133785BS							
Oil and Grease	mg/L	20.0	16.8		84	1	78-114	12/06/2024

Laboratory Control Sample Summary

Client: VERINA CONSULTING GROUP, LLC

SDG No.: P5141

Project: Rotor Clip - PO# 5183.0001

Run No.: LB133837

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133837BS							
COD	mg/L	50	49.3		99	1	90-110	12/09/2024

Laboratory Control Sample Summary

Client: VERINA CONSULTING GROUP, LLC

SDG No.: P5141

Project: Rotor Clip - PO# 5183.0001

Run No.: LB133838

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133838BS							
TSS	mg/L	550	530		96	1	90-110	12/09/2024

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	P5141
Project:	Rotor Clip - PO# 5183.0001	Run No.:	LB133842

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB165461BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	12/09/2024

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	P5141
Project:	Rotor Clip - PO# 5183.0001	Run No.:	LB133847

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB165498BS							
Cyanide	mg/L	0.1	0.10		100	1	85-115	12/09/2024