

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133866

Review By	mohan	Review On	12/13/2024 3:24:42 AM
Supervise By	jaswal	Supervise On	12/13/2024 3:25:04 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558		
ICV Standard	MP83559		
CCV Standard	MP83562		
ICSA Standard	MP83560,MP83561		
CRI Standard	MP83558		
LCS Standard			
Chk Standard	MP83565,MP83566		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	12/10/24 10:38		Kareem	OK
2	S1	S1	CAL2	12/10/24 10:43		Kareem	OK
3	S2	S2	CAL3	12/10/24 10:47		Kareem	OK
4	S3	S3	CAL4	12/10/24 10:51		Kareem	OK
5	S4	S4	CAL5	12/10/24 10:56		Kareem	OK
6	S5	S5	CAL6	12/10/24 11:00		Kareem	OK
7	ICV01	ICV01	ICV	12/10/24 11:04		Kareem	OK
8	LLICV01	LLICV01	LLICV	12/10/24 11:08		Kareem	OK
9	ICB01	ICB01	ICB	12/10/24 11:20		Kareem	OK
10	CRI01	CRI01	CRDL	12/10/24 11:24		Kareem	OK
11	ICSA01	ICSA01	ICSA	12/10/24 11:36		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	12/10/24 11:40		Kareem	OK
13	CCV01	CCV01	CCV	12/10/24 11:44		Kareem	OK
14	CCB01	CCB01	CCB	12/10/24 11:48		Kareem	OK
15	PB165463BL	PB165463BL	MB	12/10/24 11:53		Kareem	OK
16	PB165463BS	PB165463BS	LCS	12/10/24 11:57	0.1ml of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
17	PB165464BL	PB165464BL	MB	12/10/24 12:01		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133866

Review By	mohan	Review On	12/13/2024 3:24:42 AM
Supervise By	jaswal	Supervise On	12/13/2024 3:25:04 AM

STD. NAME	STD REF.#
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558
ICV Standard	MP83559
CCV Standard	MP83562
ICSA Standard	MP83560,MP83561
CRI Standard	MP83558
LCS Standard	
Chk Standard	MP83565,MP83566

18	PB165464BS	PB165464BS	LCS	12/10/24 12:06	0.1ml of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
19	P5145-01	286085	SAM	12/10/24 12:10		Kareem	OK
20	P5074-02	COMP	SAM	12/10/24 12:14		Kareem	OK
21	P5093-01	LL-001	SAM	12/10/24 12:18		Kareem	OK
22	P5093-02	LL-001-FB-12-4-24	SAM	12/10/24 12:22		Kareem	OK
23	P5093-02DUP	LL-001-FB-12-4-24DUP	DUP	12/10/24 12:27		Kareem	OK
24	P5093-02L	LL-001-FB-12-4-24L	SD	12/10/24 12:31		Kareem	OK
25	CCV02	CCV02	CCV	12/10/24 12:35		Kareem	OK
26	CCB02	CCB02	CCB	12/10/24 12:39		Kareem	OK
27	P5093-02MS	LL-001-FB-12-4-24MS	MS	12/10/24 12:44	0.1ml of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
28	P5093-02MSD	LL-001-FB-12-4-24MSD	MSD	12/10/24 12:48	0.1ml of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
29	P5093-02A	LL-001-FB-12-4-24A	PS	12/10/24 12:53	0.1ml of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
30	P5141-01	WATER TREATMENT	SAM	12/10/24 12:57		Kareem	OK
31	P5174-01	ROLL-OFF-COMP	SAM	12/10/24 13:01		Kareem	OK
32	P5175-01	OK-01-12062024	SAM	12/10/24 13:05		Kareem	OK
33	P5176-01	EO-02-12062024	SAM	12/10/24 13:09		Kareem	OK
34	P5176-03	EO-03-12062024	SAM	12/10/24 13:13		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133866

Review By	mohan	Review On	12/13/2024 3:24:42 AM
Supervise By	jaswal	Supervise On	12/13/2024 3:25:04 AM

STD. NAME	STD REF.#
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558
ICV Standard	MP83559
CCV Standard	MP83562
ICSA Standard	MP83560,MP83561
CRI Standard	MP83558
LCS Standard	
Chk Standard	MP83565,MP83566

35	P5196-01	MH-761	SAM	12/10/24 13:18		Kareem	OK
36	P5196-01DUP	MH-761DUP	DUP	12/10/24 13:22		Kareem	OK
37	CCV03	CCV03	CCV	12/10/24 13:26		Kareem	OK
38	CCB03	CCB03	CCB	12/10/24 13:30		Kareem	OK
39	P5196-01L	MH-761L	SD	12/10/24 13:34		Kareem	OK
40	P5196-01MS	MH-761MS	MS	12/10/24 13:39	0.1ml of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
41	P5196-01MSD	MH-761MSD	MSD	12/10/24 13:43	0.1ml of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
42	P5196-01A	MH-761A	PS	12/10/24 13:47	0.1ml of m6001 and m6010 were added to 10ml of the sample	Kareem	OK
43	LR1	LR1	HIGH STD	12/10/24 13:52		Kareem	OK
44	LR2	LR2	HIGH STD	12/10/24 13:57		Kareem	OK
45	CCV04	CCV04	CCV	12/10/24 14:19		Kareem	OK
46	CCB04	CCB04	CCB	12/10/24 14:24		Kareem	OK