

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** P5141
Contract: VERI01 **Lab Code:** CHEM **Case No.:** P5141 **SAS No.:** P5141
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	3.45	1.0	345	-5	7	12/10/2024	11:36	LB133866
	Chromium	56.5	52.0	109	42	62	12/10/2024	11:36	LB133866
	Copper	5.99	2.0	300	-18	22	12/10/2024	11:36	LB133866
	Lead	10.4			-12	12	12/10/2024	11:36	LB133866
	Nickel	31.3	2.0	1565	-38	42	12/10/2024	11:36	LB133866
	Silver	-2.00			-10	10	12/10/2024	11:36	LB133866
	Zinc	-6.01			-40	40	12/10/2024	11:36	LB133866
ICSAB01	Cadmium	952	972	98	826	1120	12/10/2024	11:40	LB133866
	Chromium	539	542	99	460	624	12/10/2024	11:40	LB133866
	Copper	470	511	92	434	588	12/10/2024	11:40	LB133866
	Lead	59.2	49.0	121	37	61	12/10/2024	11:40	LB133866
	Nickel	974	954	102	810	1100	12/10/2024	11:40	LB133866
	Silver	190	201	94	170	232	12/10/2024	11:40	LB133866
	Zinc	911	952	96	809	1095	12/10/2024	11:40	LB133866