

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC

SDG No.: P5141

Contract: VERI01

Lab Code: CHEM

Case No.: P5141

SAS No.: P5141

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165464BS							
Cadmium	ug/L	100	95.5		96	80 - 120	P
Chromium	ug/L	200	205		102	80 - 120	P
Copper	ug/L	150	156		104	80 - 120	P
Lead	ug/L	500	480		96	80 - 120	P
Nickel	ug/L	250	245		98	80 - 120	P
Silver	ug/L	37.5	37.7		100	80 - 120	P
Zinc	ug/L	100	106		106	80 - 120	P