

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: P5143

Project: Linden Generating Station - LRSA

Run No.: LB133770

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133770BS							
BOD5	mg/L	198	212		107	1	84.6-115.4	12/05/2024

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: P5143

Project: Linden Generating Station - LRSA

Run No.: LB133837

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133837BS							
COD	mg/L	50	49.3		99	1	90-110	12/09/2024

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: P5143

Project: Linden Generating Station - LRSA

Run No.: LB133838

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133838BS							
TSS	mg/L	550	530		96	1	90-110	12/09/2024