



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: P5146

Project: Pre Treatment Plant

RunNo.: LB133795

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Orthophosphate as P	mg/L	0.501	0.50	100	90-110	12/06/2024
Sample ID: CCV2 Orthophosphate as P	mg/L	0.501	0.5	100	90-110	12/06/2024
Sample ID: CCV1 Orthophosphate as P	mg/L	0.502	0.5	100	90-110	12/09/2024

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: P5146

Project: Pre Treatment Plant

RunNo.: LB133797

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.494	0.50	99	90-110	12/06/2024
Sample ID: CCV1 Phosphorus, Total	mg/L	0.491	0.50	98	90-110	12/06/2024
Sample ID: CCV2 Phosphorus, Total	mg/L	0.491	0.50	98	90-110	12/06/2024

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: P5146

Project: Pre Treatment Plant

RunNo.: LB133842

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	12/09/2024
Sample ID: CCV1 Ammonia as N	mg/L	1.1	1	110	90-110	12/09/2024
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	12/09/2024
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	12/09/2024
Sample ID: CCV4 Ammonia as N	mg/L	1	1	100	90-110	12/09/2024