

Initial and Continuing Calibration Blank Summary

Client: Holland Manufacturing Co.

SDG No.: P5146

Project: Pre Treatment Plant

RunNo.: LB133795

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Orthophosphate as P	mg/L	< 0.0250	0.0250	U	0.0040	0.05	12/06/2024
Sample ID: CCB1 Orthophosphate as P	mg/L	< 0.0250	0.0250	U	0.0040	0.05	12/06/2024
Sample ID: CCB2 Orthophosphate as P	mg/L	< 0.0250	0.0250	U	0.0040	0.05	12/06/2024

Initial and Continuing Calibration Blank Summary

Client: Holland Manufacturing Co.

SDG No.: P5146

Project: Pre Treatment Plant

RunNo.: LB133797

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Phosphorus, Total	mg/L	0.008	0.0250	J	0.0047	0.05	12/06/2024
Sample ID: CCB1 Phosphorus, Total	mg/L	0.008	0.0250	J	0.0047	0.05	12/06/2024
Sample ID: CCB2 Phosphorus, Total	mg/L	0.008	0.0250	J	0.0047	0.05	12/06/2024

Initial and Continuing Calibration Blank Summary

Client: Holland Manufacturing Co.

SDG No.: P5146

Project: Pre Treatment Plant

RunNo.: LB133842

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/09/2024
Sample ID: CCB1 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/09/2024
Sample ID: CCB2 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/09/2024
Sample ID: CCB3 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/09/2024
Sample ID: CCB4 Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/09/2024