

Preparation Blank Summary

Client: Holland Manufacturing Co.

SDG No.: P5146

Project: Pre Treatment Plant

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB133770BL BOD5	mg/L	< 0.2000	0.2000	U	0.17	2.0	12/05/2024
Sample ID: LB133785BL Oil and Grease	mg/L	< 2.5000	2.5000	U	0.4	5.0	12/06/2024
Sample ID: LB133795BL Orthophosphate as P	mg/L	< 0.0250	0.0250	U	0.004	0.05	12/06/2024
Sample ID: LB133838BL TSS	mg/L	< 2.0000	2.0000	U	1	4	12/09/2024
Sample ID: PB165447BL Phosphorus, Total	mg/L	0.008	0.0250	J	0.005	0.05	12/06/2024
Sample ID: PB165461BL Ammonia as N	mg/L	< 0.0500	0.0500	U	0.045	0.1	12/09/2024