

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P5146

Project: Pre Treatment Plant

Run No.: LB133770

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133770BS							
BOD5	mg/L	198	212		107	1	84.6-115.4	12/05/2024

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P5146

Project: Pre Treatment Plant

Run No.: LB133785

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133785BS							
Oil and Grease	mg/L	20.0	16.8		84	1	78-114	12/06/2024

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P5146

Project: Pre Treatment Plant

Run No.: LB133795

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133795BS							
Orthophosphate as P	mg/L	0.5	0.50		100	1	90-110	12/06/2024

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P5146

Project: Pre Treatment Plant

Run No.: LB133838

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133838BS							
TSS	mg/L	550	530		96	1	90-110	12/09/2024

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P5146

Project: Pre Treatment Plant

Run No.: LB133797

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB165447BS							
Phosphorus, Total	mg/L	0.50	0.49		98	1	90-110	12/06/2024

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: P5146

Project: Pre Treatment Plant

Run No.: LB133842

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB165461BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	12/09/2024