

Continuing Calibration Report for SequenceID : FC120624AL

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FC067927.D

|                   |              |         |        |        |            |            |        |
|-------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aliphatic C9-C12  | 10092031.000 | 60.000  | 3.154  | 6.430  | 168200.517 | 165791.359 | -1.453 |
| Aliphatic C12-C16 | 6852156.000  | 40.000  | 6.431  | 9.821  | 171303.900 | 170384.422 | -0.540 |
| Aliphatic C16-C21 | 9642607.000  | 60.000  | 9.822  | 13.180 | 160710.117 | 169621.553 | 5.254  |
| Aliphatic C21-C28 | 11401511.000 | 80.000  | 13.181 | 16.835 | 142518.888 | 157280.060 | 9.385  |
| Aliphatic C28-C40 | 13493751.000 | 120.000 | 16.836 | 21.671 | 112447.925 | 121111.336 | 7.153  |
| Aliphatic EPH     | 51482056.000 | 360.000 | 3.154  | 21.671 | 143005.711 | 150155.324 | 4.761  |