

Continuing Calibration Report for SequenceID : FC120624AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067937.D

Aliphatic C9-C12	9643965.000	60.000	3.154	6.430	160732.750	165791.359	3.051
Aliphatic C12-C16	6626528.000	40.000	6.431	9.821	165663.200	170384.422	2.771
Aliphatic C16-C21	9593540.000	60.000	9.822	13.180	159892.333	169621.553	5.736
Aliphatic C21-C28	11578614.000	80.000	13.181	16.835	144732.675	157280.060	7.978
Aliphatic C28-C40	13826870.000	120.000	16.836	21.671	115223.917	121111.336	4.861
Aliphatic EPH	51269517.000	360.000	3.154	21.671	142415.325	150155.324	5.155