

Continuing Calibration Report for SequenceID : FC120924AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067940.D

Aliphatic C9-C12	10666239.000	60.000	3.152	6.430	177770.650	165791.359	-7.226
Aliphatic C12-C16	7362587.000	40.000	6.431	9.821	184064.675	170384.422	-8.029
Aliphatic C16-C21	10474922.000	60.000	9.822	13.180	174582.033	169621.553	-2.924
Aliphatic C21-C28	12638384.000	80.000	13.181	16.834	157979.800	157280.060	-0.445
Aliphatic C28-C40	15037322.000	120.000	16.835	21.670	125311.017	121111.336	-3.468
Aliphatic EPH	56179454.000	360.000	3.152	21.670	156054.039	150155.324	-3.928