

Continuing Calibration Report for SequenceID : FC120924AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067945.D

Aliphatic C9-C12	11103446.000	60.000	3.152	6.430	185057.433	165791.359	-11.621
Aliphatic C12-C16	7622214.000	40.000	6.431	9.821	190555.350	170384.422	-11.838
Aliphatic C16-C21	10865704.000	60.000	9.822	13.180	181095.067	169621.553	-6.764
Aliphatic C21-C28	13029285.000	80.000	13.181	16.834	162866.063	157280.060	-3.552
Aliphatic C28-C40	15636843.000	120.000	16.835	21.670	130307.025	121111.336	-7.593
Aliphatic EPH	58257492.000	360.000	3.152	21.670	161826.367	150155.324	-7.773