

Continuing Calibration Report for SequenceID : FD120624AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048835.D

Aromatic C10-C12	7755997.000	40.000	4.086	5.803	193899.925	192017.544	-0.980
Aromatic C12-C16	11742896.000	60.000	5.804	8.408	195714.933	192458.977	-1.692
Aromatic C16-C21	14143269.000	80.000	8.409	12.672	176790.863	185804.816	4.851
Aromatic C21-C36	26842313.000	180.000	12.673	18.079	149123.961	157231.056	5.156
Aromatic EPH	60484475.000	360.000	4.086	18.079	168012.431	173317.266	3.061