

Continuing Calibration Report for SequenceID : FD120924AR

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FD048848.D

|                  |              |         |        |        |            |            |       |
|------------------|--------------|---------|--------|--------|------------|------------|-------|
| Aromatic C10-C12 | 7506036.000  | 40.000  | 4.086  | 5.802  | 187650.900 | 192017.544 | 2.274 |
| Aromatic C12-C16 | 11075044.000 | 60.000  | 5.803  | 8.408  | 184584.067 | 192458.977 | 4.092 |
| Aromatic C16-C21 | 13685145.000 | 80.000  | 8.409  | 12.670 | 171064.313 | 185804.816 | 7.933 |
| Aromatic C21-C36 | 25563417.000 | 180.000 | 12.671 | 18.079 | 142018.983 | 157231.056 | 9.675 |
| Aromatic EPH     | 57829642.000 | 360.000 | 4.086  | 18.079 | 160637.894 | 173317.266 | 7.316 |