

Continuing Calibration Report for SequenceID : FD120924AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048852.D

Aromatic C10-C12	7675345.000	40.000	4.086	5.802	191883.625	192017.544	0.070
Aromatic C12-C16	11281054.000	60.000	5.803	8.408	188017.567	192458.977	2.308
Aromatic C16-C21	13808809.000	80.000	8.409	12.670	172610.113	185804.816	7.101
Aromatic C21-C36	25742397.000	180.000	12.671	18.079	143013.317	157231.056	9.043
Aromatic EPH	58507605.000	360.000	4.086	18.079	162521.125	173317.266	6.229