

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** ENTACT
Lab Code: CHEM **Cas No:** P5147 **SAS No :** P5147 **SDG No:** P5147
Sample No : PB165427BSD **Datafile:** FC067930.D
Client ID : PB165427BSD

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aliphatic C9-C12	10	4.30	43		14.8 (40-140)	50
Aliphatic C12-C16	6.7	4.24	64		10.6 (40-140)	50
Aliphatic C16-C21	10	6.79	68		9.2 (40-140)	50
Aliphatic C21-C28	13.3	8.85	66		10.9 (40-140)	50
Aliphatic C28-C40	20.0	15.1	76		12.9 (40-140)	50