

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5150

Client: GFE LLC

Analytical Method: SWTO-15

Datafile : VL041736.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VL1211ABS02	Vinyl Chloride	10	12.3	ppbv	123			70	130	
	Heptane	10	9.30	ppbv	93			70	130	
	1,1-Dichloroethene	10	11.5	ppbv	115			70	130	
	Cyclohexane	10	9.20	ppbv	92			70	130	
	cis-1,2-Dichloroethene	10	9.20	ppbv	92			70	130	
	1,1,1-Trichloroethane	10	9.30	ppbv	93			70	130	
	2,2,4-Trimethylpentane	10	9.40	ppbv	94			70	130	
	Benzene	10	9.40	ppbv	94			70	130	
	1,2-Dichloroethane	10	9.60	ppbv	96			70	130	
	Trichloroethene	10	13.0	ppbv	130			70	130	
	Toluene	10	9.20	ppbv	92			70	130	
	Tetrachloroethene	10	8.90	ppbv	89			70	130	
	Ethyl Benzene	10	9.30	ppbv	93			70	130	
	m/p-Xylene	20	18.5	ppbv	93			70	130	
	o-Xylene	10	9.40	ppbv	94			70	130	
	1,3,5-Trimethylbenzene	10	9.60	ppbv	96			70	130	
	1,2,4-Trimethylbenzene	10	9.70	ppbv	97			70	130	
	Naphthalene	10	11.1	ppbv	111			70	130	
	Hexane	10	9.20	ppbv	92			70	130	