

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51847 MA No.: SDG No.: ME28N9
Instrument ID: CV1 Start Date: 12/12/2024
Analytical Method: CVAA Run Batch: LB133907
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	0	0.055	0	0.2	0.21	5	2.5	2.5	-2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51847 MA No.: _____ SDG No.: ME28N9
Instrument ID: CV1 Start Date: 12/12/2024
Analytical Method: CVAA Run Batch: LB133907
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	5	5.0	-1	7.5	7.5	0	10	10	1