
From: Panday, Prakriti P <Prakriti.Panday@gdit.com>
Sent: Monday, December 16, 2024 1:16 PM
To: Sohil Jodhani; Nimisha Pandya; Mohammad Ahmed
Cc: R5 - R5RSCC@epa.gov (R5RSCC@epa.gov)
Subject: REGION 5 | CASE 51847 | LAB ACE | SDGs E28P5 and E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT | FINAL
Attachments: TR COC.pdf

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good afternoon,

Please refer to the Region's response below regarding the sample size and QC requirement for Case 51847/SDGs E28P5 & E28Q6. Also, please be sure to provide detailed information about these issues in the SDG Narrative and let me know if you have any questions.

LAB: "Lab is sending this email with regards to case 51847.

Lab has received soil samples E28W5 & E28W6 for SVOA, SVOA-SIM and PEST analysis. These samples expected to have very high concentration of target analytes as having strong gasoline order. During extraction of Pesticides and SVOA, sample extract was very viscous and couldn't filter as well therefore lab would like to confirm that lab needs to use 1g sample instead of 30g sample due to sample matrix for Pest and SVOA, SVOA-SIM analysis. Lab QC is required for SVOA and SIM analysis and samples we are extracting with medium level extraction therefore Lab QC can not be performed for these samples due to high matrix interference therefore lab would like to confirm that the SEDD defects associated with medium level SVOA & SIM analysis should be invalid as well.

Please see attached."

REGION: "Per the client, the lab's proposal to proceed with smaller sample size is acceptable... Confirmed that the defect is invalid."

QSS INPUT: Based on the Region's response included in the email trail, the SEDD applicable defects associated with the medium level SVOA and SVOA- SIM analysis will be removed for SDGs E28P5 and E28Q6 during the CCS screening process.

***** Please note that in order to waive/remove defects associated to this issue, please send a detailed waiver request to the appropriate ASB PM.

Thank you,

PRAKRITI PANDAY
Associate Environmental Analyst
571-533-3261
GDIT Federal Civilian Division
Under Contract to EPA

15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

Leave Alert: January 17th, 2025

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: R5RSCC <R5RSCC@epa.gov>

Sent: Monday, December 16, 2024 11:58 AM

To: Panday, Prakriti P <Prakriti.Panday@gdit.com>

Subject: RE: REGION 5 | CASE 51847 | LAB ACE | SDG E28P5 & E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hi Prakriti,

Confirmed that the defect is invalid.

Thanks,
Amanda

Amanda Wroble, Ph.D.
Metals Chemist & Regional Sample Control Coordinator (RSCC)/CRL Sample Coordinator
USEPA Region 5 LSASD/AIMB
Chicago Regional Laboratory
536 S. Clark St. LAB-10C, 10th Floor
Chicago, IL 60605
(312) 353-0375

From: Panday, Prakriti P <Prakriti.Panday@gdit.com>

Sent: Monday, December 16, 2024 10:52 AM

To: R5RSCC <R5RSCC@epa.gov>

Subject: REGION 5 | CASE 51847 | LAB ACE | SDG E28P5 & E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Good morning Region 5,

As a follow up to the laboratory's inquiry below involving Case 51847/SDGs E28P5 and E28Q6, please advise if the SVOA and SVOA-SIM laboratory QC requirement is required for the medium level analysis.

LAB: "Lab is sending this email with regards to case 51847.

Lab has received soil samples E28W5 & E28W6 for SVOA, SVOA-SIM and PEST analysis. These samples expected to have very high concentration of target analytes as having strong gasoline order. During extraction of Pesticides and SVOA, sample extract was very viscous and couldn't filter as well therefore lab would like to confirm that lab needs to use 1g sample instead of 30g sample due to sample matrix for Pest and SVOA, SVOA-SIM analysis. Lab QC is required for SVOA and SIM analysis and samples we are extracting with medium level extraction therefore Lab QC can not be performed for these samples due to high matrix interference therefore lab would like to confirm that the SEDD defects associated with medium level SVOA & SIM analysis should be invalid as well.

Please see attached."

REGION: "Per the client, the lab's proposal to proceed with smaller sample size is acceptable."

***** Please note that in order to waive/remove defects associated to this issue, please send a detailed waiver request to the appropriate ASB PM.

Thank you,

PRAKRITI PANDAY
Associate Environmental Analyst
571-533-3261
GDIT Federal Civilian Division
Under Contract to EPA
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

Leave Alert: January 17th, 2025

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: R5RSCC <R5RSCC@epa.gov>
Sent: Monday, December 16, 2024 8:36 AM
To: Panday, Prakriti P <Prakriti.Panday@gdit.com>
Subject: FW: REGION 5 | CASE 51847 | LAB ACE | SDG E28P5 & E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hi Prakriti,

Per the client, the lab's proposal to proceed with smaller sample size is acceptable.

Thanks,
Amanda

Amanda Wroble, Ph.D.
Metals Chemist & Regional Sample Control Coordinator (RSCC)/CRL Sample Coordinator
USEPA Region 5 LSASD/AIMB
Chicago Regional Laboratory
536 S. Clark St. LAB-10C, 10th Floor
Chicago, IL 60605
(312) 353-0375

From: Hares, Sarah M CIV USARMY CELRB (USA) <Sarah.M.Hares@usace.army.mil>
Sent: Friday, December 13, 2024 12:18 PM
To: R5RSCC <R5RSCC@epa.gov>
Cc: Green, Nilia <Green.Nilia@epa.gov>; Kowal, Shawn F CIV USARMY CELRB (USA) <Shawn.F.Kowal@usace.army.mil>
Subject: RE: REGION 5 | CASE 51847 | LAB ACE | SDG E28P5 & E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Hi Amanda,

Our recommendation is to follow the lab's standard procedure for samples with matrix interference (e.g., dilution or smaller sample size). It wasn't explicitly clear on how to proceed after reviewing the CLP SOW.

Thank you,

Sarah Hares
Environmental Engineer
U.S. Army Corps of Engineers
Buffalo District

From: R5RSCC <R5RSCC@epa.gov>
Sent: Thursday, December 12, 2024 8:27 AM
To: Hares, Sarah M CIV USARMY CELRB (USA) <Sarah.M.Hares@usace.army.mil>
Cc: Green, Nilia <Green.Nilia@epa.gov>; Kowal, Shawn F CIV USARMY CELRB (USA) <Shawn.F.Kowal@usace.army.mil>
Subject: [Non-DoD Source] FW: REGION 5 | CASE 51847 | LAB ACE | SDG E28P5 & E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Hi Sarah,

Please advise on the issue below.

Thanks,
Amanda

Amanda Wroble, Ph.D.
Metals Chemist & Regional Sample Control Coordinator (RSCC)/CRL Sample Coordinator
USEPA Region 5 LSASD/AIMB

Chicago Regional Laboratory
536 S. Clark St. LAB-10C, 10th Floor
Chicago, IL 60605
(312) 353-0375

From: Panday, Prakriti P <Prakriti.Panday@gdit.com>

Sent: Wednesday, December 11, 2024 4:08 PM

To: R5RSCC <R5RSCC@epa.gov>

Subject: REGION 5 | CASE 51847 | LAB ACE | SDG E28P5 & E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Good afternoon,

Please advise on the laboratory's reported issue for Case 51847/SDGs E28P5 and E28Q6.

LAB: "Lab is sending this email with regards to case 51847.

Lab has received soil samples E28W5 & E28W6 for SVOA, SVOA-SIM and PEST analysis. These samples expected to have very high concentration of target analytes as having strong gasoline order. During extraction of Pesticides and SVOA, sample extract was very viscous and couldn't filter as well therefore lab would like to confirm that lab needs to use 1g sample instead of 30g sample due to sample matrix for Pest and SVOA, SVOA-SIM analysis. Lab QC is required for SVOA and SIM analysis and samples we are extracting with medium level extraction therefore Lab QC can not be performed for these samples due to high matrix interference therefore lab would like to confirm that the SEDD defects associated with medium level SVOA & SIM analysis should be invalid as well.

Please see attached."

Thank you,

PRAKRITI PANDAY

Associate Environmental Analyst

571-533-3261

GDIT Federal Civilian Division

Under Contract to EPA

15036 Conference Center Drive

Chantilly, VA 20151

www.gdit.com

GENERAL DYNAMICS
Information Technology

Leave Alert: January 17th, 2025

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: DASSupport

Sent: Wednesday, December 11, 2024 5:01 PM

To: 'Sohil Jodhani' <Sohil.Jodhani@alliancetg.com>

Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>

Subject: REGION 5 | CASE 51847 | LAB ACE | SDG E28P5 & E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Thank you for contacting DAS Support. This is to confirm that your email has been received and is currently being addressed. An updated response will be provided as soon as it is available.

Thank you,

The DAS Support Team
DASsupport@gdit.com

GDIT Federal Civilian Division
Under Contract to EPA
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Sohil Jodhani <Sohil.Jodhani@alliancetg.com>

Sent: Wednesday, December 11, 2024 3:53 PM

To: DASsupport <DASsupport@gdit.com>

Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>

Subject: REGION 5 | CASE 51847 | LAB ACE | SDG E28P5 & E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT |

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hello,

Lab is sending this email with regards to case 51847.

Lab has received soil samples E28W5 & E28W6 for SVOA, SVOA-SIM and PEST analysis. These samples expected to have very high concentration of target analytes as having strong gasoline order. During extraction of Pesticides and SVOA, sample extract was very viscous and couldn't filter as well therefore lab would like to confirm that lab needs to use 1g sample instead of 30g sample due to sample matrix for Pest and SVOA, SVOA-SIM analysis. Lab QC is required for SVOA and SIM analysis and samples we are extracting with medium level extraction therefore Lab QC can not be performed for these samples due to high matrix

interference therefore lab would like to confirm that the SEDD defects associated with medium level SVOA & SIM analysis should be invalid as well.

Please see attached.

Thanks & Regards,



Sohil Jodhani

QA/QC Director

An Alliance Technical Group Company

Main: 908-789-8900

Direct: 908-728-3152

Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092

www.alliancetg.com



From: Sohil Jodhani
Sent: Friday, December 27, 2024 9:54 AM
To: Deepak Parmar
Subject: Fw: REGION 5 | CASE 51847 | LAB ACE | SDGs E28R5 & E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT | FINAL
Attachments: cccheck.pdf; P5232-04(E28R8)B.pdf; P5155-15(E28W0)B.pdf; P5155-14(E28T9)B.pdf; P5155-12(E28T7)B.pdf; P5155-10(E28T5)B.pdf; P5155-09(E28T4)B.pdf; P5232-08(E28S2)B.pdf; P5232-06(E28S0)B.pdf; P5155-07(E28R2)B.pdf; P5155-06(E28R1)B.pdf; P5232-01(E28R5)B.pdf; P5155-05(E28R0)B.pdf

Please save in PM folder.

Thanks and Regards

Sohil Jodhani
QA/QC Director
Chemtech
An Alliance Technical Group Company
908-728-3152

From: Irving, Madison <Madison.Irving@gdit.com>
Sent: Friday, December 27, 2024 9:39:46 AM
To: Sohil Jodhani <Sohil.Jodhani@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>; Mohammad Ahmed <mohammad.ahmed@alliancetg.com>
Cc: R5 - R5RSCC@epa.gov (R5RSCC@epa.gov) <R5RSCC@epa.gov>
Subject: REGION 5 | CASE 51847 | LAB ACE | SDGs E28R5 & E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT | FINAL

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good morning ACE,

Regarding the laboratory's reported issue for Case 51847/SDG E28R5 & E28Q6, please see the Region and QSS input below and let me know if you have any questions.

LAB: "Lab has received soil samples for VOA analysis. Lab has performed few samples reanalysis for low level VOA analysis and closing CCV has more than two analytes outside the QC limits. Since, these were the re-analysis therefore Lab will report both original and re-analysis with closing CCV failure in final electronic deliverables.

Please see attachment for your reference."

QSS INPUT: The laboratory is expected to comply with the SOW requirements specified in Exhibit D, Sections 9.5.6 and 11.4.

REGION: "The client accepts the lab's proposal of reporting both sets of data."

***** Please note that in order to waive/remove defects associated to this issue, please send a detailed waiver request to the appropriate ASB PM.

Thank you,

Madison Irving
Associate Environmental Analyst
GDIT Federal Civilian Division
Under Contract to EPA
T: 919-768-4149
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com



LEAVE ALERT: None

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: R5RSCC <R5RSCC@epa.gov>
Sent: Friday, December 27, 2024 9:17 AM
To: Danganan, Eloise B <Eloise.Danganan@gdit.com>
Cc: Irving, Madison <Madison.Irving@gdit.com>
Subject: RE: REGION 5 | CASE 51847 | LAB ACE | SDGs E28R5 & E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DASS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Good morning,

Sorry for the delayed response. I was waiting to hear back from the client.

The client accepts the lab's proposal of reporting both sets of data.

Thanks,

Darcie D. Tone-Pah-Hote

Regional Sample Control Coordinator

LSASD-SQAB
U.S. EPA Region 5
77 W Jackson Blvd
Chicago, IL 60604
Cubicle #: 06078
Phone #: 312-886-4298

From: Danganan, Eloise B <Eloise.Danganan@gdit.com>
Sent: Tuesday, December 24, 2024 9:48 AM
To: R5RSCC <R5RSCC@epa.gov>; Tone-Pah-Hote, Darcie <TonePahHote.Darcie@epa.gov>
Subject: RE: REGION 5 | CASE 51847 | LAB ACE | SDGs E28R5 & E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DASS SUPPORT

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Good morning Region 5,

Please advise on the issue below involving Case 51847/SDGs E28R5 and E28Q6. If a response was already sent to Madison Irving, please forward it to me. She is out of the office and I do not have access to her email.

Thank you,

ELOISE DANGANAN
Environmental Analyst
GDIT Federal Civilian Division
Under contract to the EPA
T 703.254.0067
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com



LEAVE ALERT: none

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Irving, Madison <Madison.Irving@gdit.com>
Sent: Friday, December 20, 2024 1:11 PM
To: R5 - R5RSCC@epa.gov (<R5RSCC@epa.gov>) <R5RSCC@epa.gov>
Subject: REGION 5 | CASE 51847 | LAB ACE | SDGs E28R5 & E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DASS SUPPORT

Good afternoon Region 5,

Please advise on the laboratory's reported issue below for Case 51847/SDG E28R5 & E28Q6.

LAB: "Lab has received soil samples for VOA analysis. Lab has performed few samples reanalysis for low level VOA analysis and closing CCV has more than two analytes outside the QC limits. Since, these were the re-analysis therefore Lab will report both original and re-analysis with closing CCV failure in final electronic deliverables.

Please see attachment for your reference."

QSS Input:

The laboratory is expected to comply with the SOW requirements specified in Exhibit -D-Sections, 9.5.6 and 11.4.

Thank you,

Madison Irving
Associate Environmental Analyst
GDIT Federal Civilian Division
Under Contract to EPA
T: 919-768-4149
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com



LEAVE ALERT: None

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: DASsupport
Sent: Friday, December 20, 2024 11:50 AM
To: Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>
Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>
Subject: REGION 5 | CASE 51847 | LAB ACE | SDGs E28R5 & E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DASS SUPPORT

Thank you for contacting DAS Support. This is to confirm that your email has been received and is currently being addressed. An updated response will be provided as soon as it is available.

Thank you,

The DAS Support Team
DASsupport@gdit.com

GDIT Federal Civilian Division
Under contract to the EPA
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>
Sent: Friday, December 20, 2024 11:38 AM
To: DASsupport <DASsupport@gdit.com>
Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>
Subject: REGION 5 | CASE 51847 | LAB ACE | SDGs E28R5 & E28Q6 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DASS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hello,

Lab is sending this email with regards to case 51847.

Lab has received soil samples for VOA analysis. Lab has performed few samples reanalysis for low level VOA analysis and closing CCV has more than two analytes outside the QC limits. Since, these were the re-analysis therefore Lab will report both original and re-analysis with closing CCV failure in final electronic deliverables.

Please see attachment for your reference.

Thanks & Regards,



Sohil Jodhani
QA/QC Director
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3152
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com     

From: Maheshbhai Dadoda <Maheshbhai.Dadoda@alliancetg.com>
Sent: Friday, December 20, 2024 11:17 AM
To: Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>
Subject: VD121224-SDG 5232,5155

Sequence has endcheck standard failure(more than 2 failed above 50%) and we have re-analysis samples affected due to that

All analysis are ok(no internal surrogate failed no over-range analytes) except 5155-05RE, 06RE,has internal standard failure and need medium level analysis

Lab will report all analysis as confirm analysis with failed CCC standard

Thanks

Mahesh