EPA SAMPLE NO.

ME28W2S

FORM 5A - IN MATRIX SPIKE SAMPLE RECOVERY

Lab Name:	Chemtech Co	onsulting Grou	р	Contract:	EPW14030			
Lab Code:	ACE	Case No.:	51847	MA No. :		SDG No.:	ME28Q6	
Matrix :	Soil			Analytical	Method:	Spectroph	otometry	
% Solids:	75.3							
Concentrat	ion Units (μ	g/L, mg/L or m	ng/kg dry we	eight): mg/l	kg			

Analyte	Control Limit %R	Spiked Sample Result (SSR)	Q	Sample Result (SR)	Q	Spike Added (SA)	%R	Q
Cyanide	75 - 125	11		4.6		6.4	94	