FORM 15-IN

INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011

Lab Code: ACE Case No.: 51847 MA No.: _____ SDG No.: ME28Q6

Instrument ID: CV1 Start Date: 12/20/2024

Analytical Method: CVAA Run Batch: LB134038

Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	0	0.0	0	0.2	0.18	-12	2.5	2.6	2

FORM 15-IN

INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011

Lab Code: ACE Case No.: 51847 MA No.: _____ SDG No.: ME28Q6

Instrument ID: CV1 Start Date: 12/20/2024

Analytical Method: CVAA Run Batch: LB134038

Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	5	5.1	2	7.5	7.4	-1	10	10	0