



**284 Sheffield Street
Mountainside, NJ 07092**

SDG NARRATIVE

USEPA

SDG # MJNKP7

CASE # 51821

CONTRACT # 68HERH20D0011

SOW# SFAM01.1

LAB NAME: Alliance Technical Group, LLC

LAB CODE: ACE

LAB ORDER ID # P5163

A. Number of Samples and Date of Receipt

01 Soil samples was delivered to the laboratory intact on 12/06/2024

B. Parameters

Test requested for Metals CLP4 MS = Arsenic, Copper, Lead, Zinc.

C. Cooler Temp

Indicator Bottle: Presence/Absence

Cooler: 2.1°C

D. Detail Documentation (related to Sample Handling Shipping, Analytical Problem, Temp of Cooler etc):

Issue 1 : A "P" or "M" prefix was listed at the beginning of a CLP sample ID.

Issue 2: The COC indicates PRs are required for all samples, but per scheduling there are no PRs requested for this Case. Please advise on how the laboratory may proceed.

E. Corrective Action taken for above:

Resolution 1 : To maintain COC integrity, ASB requests no changes to the Sample IDs. The laboratory will note the issue in the SDG Narrative and proceed with the analysis of the samples.

Resolution 2: Per Region 10, preliminary results were inadvertently included on the COC but are not needed for this project. Please proceed with the scheduled 21-day TAT. Please note the issue in the SDG Narrative and proceed with analysis of the samples



**284 Sheffield Street
Mountainside, NJ 07092**

F. Analytical Techniques:

All analyses were based on CLP Methodology by method SFAM01.1.

G. Calculation:

Calculation for ICP-MS Soil Sample:

Conversion of Results from $\mu\text{g/L}$ or ppb to mg/kg :

$$\text{Concentration (mg/kg)} = C \times \frac{V_f}{W \times S} \times \text{DF} / 1000$$

Where,

C = Instrument value in ppb (The average of all replicate integrations)

Vf = Final digestion volume (mL)

W = Initial aliquot amount (g) (Fraction of Sample amount taken in prep)

S = % Solids / 100 (Fraction of Percent Solids)

DF = Dilution Factor

Example Calculation For Sample MJNKP7 For Arsenic:

If C = 178.64 ppb

Vf = 500 ml

W = 1.36 g

S = 0.736(73.6/100)

DF = 1

$$\text{Concentration (mg/kg)} = 178.64 \times \frac{500}{1.36 \times 0.736} \times 1 / 1000$$

$$= 89.2343 \text{ mg/kg}$$

$$= 89 \text{ mg/kg (Reported Result with Signification)}$$

H. QA/ QC

Calibrations met requirements. Interference check met requirements. Blank analyses did not indicate any presence of contamination. Laboratory Control sample was within control limits. Spike sample did meet requirements. Duplicate sample did meet requirements. Serial Dilution did meet requirements.

Collision cell is being used to remove potential interferences. The analytes Na, Mg, Al, K, V, Cr, Mn, Fe, Co, Ni, Cu, Zn, As are being analyzed with collision cell and analytes Be, B, Ca, Ti, Se, Sr, Zr, Mo, Ag, Cd, Sn, Sb, Ba, Tl, Pb, U are being analyzed with Non-Collision Cell. Helium gas is used for the Collision Cell analysis.



**284 Sheffield Street
Mountainside, NJ 07092**

Internal Standard Association for ICP-MS analysis.

Target Analyte	Associated Internal Standard
Arsenic	89Y
Copper	45Sc
Lead	209Bi
Zinc	45Sc

I certify that the data package is in compliance with the terms and conditions of the contract both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the Laboratory Director or his designee, as verified by the following signature.

Signature_____

Name: Nimisha Pandya

Date _____

Title: Document Control Officer