



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P5164

Client: AECOM Technical Services, Inc.

Analytical Method: 8270-Modified DataFile: BN035528.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB165497BS	1,4-Dioxane	0.4	0.36	ug/L	90				70	130	