

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5164 SAS No.: P5164 SDG NO.: P5164

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 12/10/2024

Lab File ID: BN035525.D Time Analyzed: 09:33

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2237	7.293	6161	10.04	4478	13.96
UPPER LIMIT	4474	7.793	12322	10.541	8956	14.457
LOWER LIMIT	1118.5	6.793	3080.5	9.541	2239	13.457
EPA SAMPLE NO.						
01 PB165497BSD	1941	7.29	5150	10.04	3668	13.96
02 PB165497BS	1669	7.30	4431	10.04	3092	13.96

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.