

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1209WBSD01		SDG No.:	P5164
Lab Sample ID:	VX1209WBSD01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044194.D	1		12/09/24 17:37	VX120924

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	16.0		0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	18.1		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	18.3		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	21.0		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	25.4		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.8		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	18.3		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.1		0.26	0.75	1.00	ug/L
67-64-1	Acetone	100		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	13.8		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	21.9		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	21.5		0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	19.9		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	19.5		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	21.2		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	18.8		1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	110		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.4		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	21.3		0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	18.9		0.18	0.50	1.00	ug/L
67-66-3	Chloroform	22.3		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	21.3		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	18.4		0.19	0.50	1.00	ug/L
71-43-2	Benzene	20.2		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	21.5		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	17.9		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	20.0		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	20.4		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.75	2.50	5.00	ug/L
108-88-3	Toluene	20.8		0.18	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1209WBSD01		SDG No.:	P5164
Lab Sample ID:	VX1209WBSD01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044194.D	1		12/09/24 17:37	VX120924

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	19.8		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	20.2		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	21.5		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	120		1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	19.8		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	21.3		0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	19.6		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	20.0		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	20.9		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	41.4		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	21.3		0.14	0.50	1.00	ug/L
100-42-5	Styrene	21.1		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	17.5		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	21.9		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	22.2		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	20.6		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	20.0		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	21.3		0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	20.7		0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	19.4		0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	20.0		0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	55.7		81 - 118		111%	SPK: 50
1868-53-7	Dibromofluoromethane	52.0		80 - 119		104%	SPK: 50
2037-26-5	Toluene-d8	50.9		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.8		85 - 114		108%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	136000	5.55				
540-36-3	1,4-Difluorobenzene	249000	6.757				
3114-55-4	Chlorobenzene-d5	216000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	96600	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX1209WBSD01		SDG No.:	P5164
Lab Sample ID:	VX1209WBSD01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044194.D	1		12/09/24 17:37	VX120924

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products