

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: P5164 SAS No.: P5164 SDG NO.: P5164
 Lab File ID: VX044197.D Date Analyzed: 12/10/2024
 Instrument ID: MSVOA_X Time Analyzed: 10:43
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	121082	5.54	212067	6.76	187191	10.05
UPPER LIMIT	242164	6.038	424134	7.257	374382	10.549
LOWER LIMIT	60541	5.038	106034	6.257	93595.5	9.549
EPA SAMPLE NO.						
TB	112437	5.55	211530	6.76	185283	10.06
VX1210WBL01	110263	5.54	207593	6.76	178508	10.06
VX1210WBS01	143461	5.54	255477	6.76	217630	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.