

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: P5164 SAS No.: P5164 SDG NO.: P5164
 Lab File ID: VX044174.D Date Analyzed: 12/09/2024
 Instrument ID: MSVOA_X Time Analyzed: 09:42
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	97559	12.018				
UPPER LIMIT	195118	12.518				
LOWER LIMIT	48779.5	11.518				
EPA SAMPLE NO.						
RE108D2-20241206	80855	12.02				
RE108D1-20241206	74622	12.02				
RE105D1-20241206	70850	12.02				
RE105D2-20241206	82399	12.02				
FB-20241206	74280	12.02				
VX1209WBL01	90543	12.02				
VX1209WBS01	95808	12.02				
VX1209WBSD01	96570	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.