

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: P5164 SAS No.: P5164 SDG NO.: P5164
 Lab File ID: VX044197.D Date Analyzed: 12/10/2024
 Instrument ID: MSVOA_X Time Analyzed: 10:43
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	91127	12.024				
UPPER LIMIT	182254	12.524				
LOWER LIMIT	45563.5	11.524				
EPA SAMPLE NO.						
TB	76755	12.02				
VX1210WBL01	71476	12.02				
VX1210WBS01	101326	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.