

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: P5164 SAS No.: P5164 SDG NO.: P5164
Lab File ID: VX044173.D BFB Injection Date: 12/09/2024
Instrument ID: MSVOA_X BFB Injection Time: 09:06
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21.5
75	30.0 - 60.0% of mass 95	57.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.6 (0.8) 1
174	50.0 - 100.0% of mass 95	69.6
175	5.0 - 9.0% of mass 174	5.4 (7.8) 1
176	95.0 - 101.0% of mass 174	68.7 (98.7) 1
177	5.0 - 9.0% of mass 176	4.3 (6.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX044174.D	12/09/2024	09:42
VX1209WBL01	VX1209WBL01	VX044176.D	12/09/2024	10:35
VX1209WBS01	VX1209WBS01	VX044178.D	12/09/2024	11:24
FB-20241206	P5164-06	VX044186.D	12/09/2024	14:30
RE108D1-20241206	P5164-02	VX044188.D	12/09/2024	15:17
RE105D1-20241206	P5164-03	VX044189.D	12/09/2024	15:40
RE108D2-20241206	P5164-01	VX044192.D	12/09/2024	16:51
RE105D2-20241206	P5164-04	VX044193.D	12/09/2024	17:14
VX1209WBSD01	VX1209WBSD01	VX044194.D	12/09/2024	17:37
VSTDCCC050EC	VSTDCCC050	VX044195.D	12/09/2024	18:00