

FORM 11-IN
ICP-MS INTERNAL STANDARD ASSOCIATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51821 MA No.: SDG No.: MJNKR9
Instrument ID: P8 Date: 12/13/2024
Run Batch: LB133950

Analyte	Assoc. Internal Standard 1	Assoc. Internal Standard 2
Aluminum	45Sc	
Antimony	159Tb	
Arsenic	89Y	
Barium	159Tb	
Beryllium	6Li	
Cadmium	159Tb	
Calcium	45Sc	
Chromium	45Sc	
Cobalt	45Sc	
Copper	45Sc	
Iron	45Sc	
Lead	209Bi	
Magnesium	45Sc	
Manganese	45Sc	
Nickel	45Sc	
Potassium	45Sc	
Selenium	89Y	
Silver	159Tb	
Sodium	45Sc	
Thallium	209Bi	
Vanadium	45Sc	
Zinc	45Sc	