

## FORM 2 - IN

## INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
 Lab Code: ACE Case No.: 51821 MA No. :                      SDG No.: MJNKR9  
 Initial Calibration Verification Source : MP83041  
 Continuing Calibration Verification Source : MP83042  
 Run Batch: LB133950 Analytical Method: ICP-MS  
 Concentration Units: µg/L

|         | Initial Calibration Verification |       |     |      | Continuing Calibration Verification |       |     |      |            |     |      |
|---------|----------------------------------|-------|-----|------|-------------------------------------|-------|-----|------|------------|-----|------|
|         | ID: ICV001                       |       |     |      | ID: CCV001                          |       |     |      | ID: CCV002 |     |      |
| Analyte | True                             | Found | %R  | %RSD | True                                | Found | %R  | %RSD | Found      | %R  | %RSD |
| Arsenic | 200                              | 220   | 109 | 1    | 500                                 | 470   | 95  | 1    | 480        | 97  | 1    |
| Copper  | 100                              | 100   | 103 | 1    | 5000                                | 4600  | 92  | 1    | 4600       | 92  | 1    |
| Lead    | 200                              | 210   | 104 | 2    | 2500                                | 2500  | 99  | 0    | 2600       | 105 | 9    |
| Zinc    | 200                              | 210   | 105 | 1    | 5000                                | 5000  | 100 | 1    | 5000       | 100 | 0    |