

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P5166 SAS No.: P5166 SDG No.: P5166

Instrument ID: BNA_N Calibration Date/Time: 12/09/2024 17:07

Lab File ID: BN035510.D Init. Calib. Date(s): 11/27/2024 11/27/2024

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 15:34 19:09

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.626	0.627		0.2	20.0
Fluoranthene-d10	1.134	1.031		-9.1	20.0
2-Fluorophenol	1.001	1.045		4.4	20.0
Phenol-d6	1.204	1.324		10.0	20.0
Nitrobenzene-d5	0.244	0.260		6.6	20.0
2-Fluorobiphenyl	1.512	1.514		0.1	20.0
2,4,6-Tribromophenol	0.284	0.274		-3.5	20.0
Terphenyl-d14	0.789	0.816		3.4	20.0
1,4-Dioxane	0.382	0.363		-5.0	20.0

All other compounds must meet a minimum RRF of 0.010.